

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014195.D
 Acq On : 23 Dec 2019 12:30
 Operator : JC/SP
 Sample : PB125652TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125652TB

Quant Time: Dec 24 07:06:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	587245	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	911265	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	815844	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	372663	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	322052	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
35) Dibromofluoromethane	5.48	113	268928	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
50) Toluene-d8	8.71	98	1082167	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
62) 4-Bromofluorobenzene	11.13	95	374597	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
Target Compounds						
20) Methylene Chloride	2.84	84	12931	1.900	ug/l #	89
43) Isopropyl Acetate	6.43	43	150672	9.860	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014195.D
Acq On : 23 Dec 2019 12:30
Operator : JC/SP
Sample : PB125652TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125652TB

Quant Time: Dec 24 07:06:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

