

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014202.D
 Acq On : 23 Dec 2019 15:13
 Operator : JC/SP
 Sample : K6393-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-BNM-1

Quant Time: Dec 24 07:40:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	576416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	888260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	806384	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	373438	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	316285	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
35) Dibromofluoromethane	5.49	113	267336	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
50) Toluene-d8	8.71	98	1051906	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.13	95	361434	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
Target Compounds						
16) Acetone	2.43	43	171180	40.361	ug/l	99
21) trans-1,2-Dichloroethene	3.15	96	16082	2.633	ug/l	89
52) Toluene	8.78	92	4879	0.308	ug/l	88
68) m/p-Xylenes	10.36	106	3792	0.334	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014202.D
Acq On : 23 Dec 2019 15:13
Operator : JC/SP
Sample : K6393-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-BNM-1

Quant Time: Dec 24 07:40:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

