

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014206.D
 Acq On : 23 Dec 2019 16:46
 Operator : JC/SP
 Sample : K6393-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-12-20-19

Quant Time: Dec 24 07:47:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	576914	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	880279	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	792442	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	350072	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	316383	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
35) Dibromofluoromethane	5.49	113	265560	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
50) Toluene-d8	8.71	98	1049959	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.13	95	356771	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
Target Compounds						
16) Acetone	2.43	43	48120	11.336	ug/l	98
20) Methylene Chloride	2.85	84	10634	1.591	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014206.D
Acq On : 23 Dec 2019 16:46
Operator : JC/SP
Sample : K6393-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-12-20-19

Quant Time: Dec 24 07:47:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

