

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014207.D
 Acq On : 23 Dec 2019 17:09
 Operator : JC/SP
 Sample : K6401-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B53

Quant Time: Dec 24 07:50:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	578091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	885499	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	791377	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	351794	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	315703	48.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.38%	
35) Dibromofluoromethane	5.49	113	264812	49.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.40%	
50) Toluene-d8	8.71	98	1039598	49.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.14	95	353698	46.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014207.D
Acq On : 23 Dec 2019 17:09
Operator : JC/SP
Sample : K6401-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B53

Quant Time: Dec 24 07:50:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

