

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014211.D
 Acq On : 23 Dec 2019 18:43
 Operator : JC/SP
 Sample : K6387-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Dec 24 08:00:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	562621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	863833	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	775631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	348612	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	309713	48.57	ug/l	0.00
Spiked Amount			Recovery	=	97.14%	
35) Dibromofluoromethane	5.49	113	255613	48.68	ug/l	0.00
Spiked Amount			Recovery	=	97.36%	
50) Toluene-d8	8.71	98	1020689	49.93	ug/l	0.00
Spiked Amount			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.14	95	348236	46.60	ug/l	0.00
Spiked Amount			Recovery	=	93.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	40259	5.648	ug/l	98
12) 1,1-Dichloroethene	2.37	96	281286	51.974	ug/l	98
24) 1,1-Dichloroethane	3.69	63	260310	24.298	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	52326	7.760	ug/l	86
40) Benzene	6.14	78	11710	0.480	ug/l #	90
42) 1,2-Dichloroethane	6.19	62	20568	2.488	ug/l	95
44) Trichloroethene	7.21	130	6214	0.922	ug/l	95
49) 1,4-Dioxane	7.73	88	75056	519.975	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014211.D
Acq On : 23 Dec 2019 18:43
Operator : JC/SP
Sample : K6387-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Dec 24 08:00:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

