

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014212.D
 Acq On : 23 Dec 2019 19:06
 Operator : JC/SP
 Sample : K6387-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Dec 24 03:32:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	558083	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	857751	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	775631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	345016	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	310461	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
35) Dibromofluoromethane	5.49	113	256276	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	8.71	98	1019593	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
62) 4-Bromofluorobenzene	11.13	95	344588	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
Target Compounds						
4) Vinyl Chloride	1.40	62	15316	2.166	ug/l	98
12) 1,1-Dichloroethene	2.37	96	53719	10.007	ug/l	99
16) Acetone	2.43	43	7607	1.852	ug/l	98
24) 1,1-Dichloroethane	3.69	63	35199	3.312	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	15628	2.336	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014212.D
Acq On : 23 Dec 2019 19:06
Operator : JC/SP
Sample : K6387-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Dec 24 03:32:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

