

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014214.D
 Acq On : 23 Dec 2019 19:52
 Operator : JC/SP
 Sample : K6387-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: Dec 24 08:08:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	559463	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	860368	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	775202	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	348558	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	312065	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
35) Dibromofluoromethane	5.49	113	260105	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
50) Toluene-d8	8.71	98	1012620	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
62) 4-Bromofluorobenzene	11.13	95	350086	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.37	96	3111	0.578	ug/l #	90
16) Acetone	2.44	43	5964	1.449	ug/l #	88
27) cis-1,2-Dichloroethene	4.59	96	3650	0.544	ug/l	72
49) 1,4-Dioxane	7.74	88	3546	24.665	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014214.D
Acq On : 23 Dec 2019 19:52
Operator : JC/SP
Sample : K6387-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: Dec 24 08:08:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

