

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014222.D
 Acq On : 24 Dec 2019 00:08
 Operator : JC/SP
 Sample : K6401-01 100X
 Misc : 4.72g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B37

Quant Time: Dec 24 08:31:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	575991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	878654	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	794960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	372883	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	316752	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
35) Dibromofluoromethane	5.49	113	262111	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	8.71	98	1049477	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	11.13	95	373372	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
Target Compounds						
44) Trichloroethene	7.21	130	23523	3.433	ug/l	97
67) Ethyl Benzene	10.25	91	231638	7.971	ug/l	100
68) m/p-Xylenes	10.36	106	316934	28.356	ug/l	99
69) o-Xylene	10.70	106	118985	10.855	ug/l	98
80) 1,3,5-Trimethylbenzene	11.50	105	60455	2.736	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	108306	4.895	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014222.D
Acq On : 24 Dec 2019 00:08
Operator : JC/SP
Sample : K6401-01 100X
Misc : 4.72g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B37

Quant Time: Dec 24 08:31:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

