

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014223.D
 Acq On : 24 Dec 2019 00:31
 Operator : JC/SP
 Sample : K6401-01 800X
 Misc : 4.72g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B37

Quant Time: Dec 24 08:33:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	576017	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	888556	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	797489	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	368304	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	319260	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
35) Dibromofluoromethane	5.49	113	262797	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
50) Toluene-d8	8.71	98	1052144	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	11.14	95	359786	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
Target Compounds						
68) m/p-Xylenes	10.36	106	30843	2.751	ug/l	99
69) o-Xylene	10.70	106	11923	1.084	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014223.D
Acq On : 24 Dec 2019 00:31
Operator : JC/SP
Sample : K6401-01 800X
Misc : 4.72g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B37

Quant Time: Dec 24 08:33:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

