

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014225.D
 Acq On : 24 Dec 2019 01:18
 Operator : JC/SP
 Sample : K6401-04 800X
 Misc : 4.58g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B38

Quant Time: Dec 24 08:38:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	570850	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	868564	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	780559	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	354813	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	309376	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
35) Dibromofluoromethane	5.49	113	257672	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
50) Toluene-d8	8.71	98	1019310	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	11.13	95	351046	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
Target Compounds						
68) m/p-Xylenes	10.36	106	21705	1.978	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014225.D
Acq On : 24 Dec 2019 01:18
Operator : JC/SP
Sample : K6401-04 800X
Misc : 4.58g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B38

Quant Time: Dec 24 08:38:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

