

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122319\
 Data File : VX014228.D
 Acq On : 24 Dec 2019 02:28
 Operator : JC/SP
 Sample : K6401-06 100X
 Misc : 5.27g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B40

Quant Time: Dec 24 04:31:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	567956	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	860589	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	784283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	364816	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	313218	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
35) Dibromofluoromethane	5.49	113	257853	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
50) Toluene-d8	8.71	98	1029062	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
62) 4-Bromofluorobenzene	11.13	95	353912	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
65) Chlorobenzene	10.14	112	16935	1.016	ug/l	95
87) 1,3-Dichlorobenzene	12.02	146	98811	7.978	ug/l	99
88) 1,4-Dichlorobenzene	12.09	146	148578	11.798	ug/l	98
91) 1,2-Dichlorobenzene	12.39	146	302178	24.236	ug/l	99
93) 1,2,4-Trichlorobenzene	13.64	180	8742	1.079	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122319\
Data File : VX014228.D
Acq On : 24 Dec 2019 02:28
Operator : JC/SP
Sample : K6401-06 100X
Misc : 5.27g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B40

Quant Time: Dec 24 04:31:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

