

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122320\
 Data File : VX020365.D
 Acq On : 24 Dec 2020 02:24
 Operator : JC/MD
 Sample : L5095-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 39-MWRE3S-121420

Quant Time: Dec 24 05:36:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	246181	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.830	114	417455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	403391	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	172756	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	168089	50.94	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.88%
35) Dibromofluoromethane	5.465	113	135339	50.68	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.36%
50) Toluene-d8	8.696	98	503809	49.01	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.02%
62) 4-Bromofluorobenzene	11.122	95	182208	46.31	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.62%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	1959	1.30	ug/l	94
21) trans-1,2-Dichloroethene	3.148	96	851	0.28	ug/l #	72
27) cis-1,2-Dichloroethene	4.562	96	47624	13.83	ug/l	88
44) Trichloroethene	7.190	130	3310	1.00	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122320\
Data File : VX020365.D
Acq On : 24 Dec 2020 02:24
Operator : JC/MD
Sample : L5095-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
39-MWRE3S-121420

Quant Time: Dec 24 05:36:04 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

