

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX122320\
 Data File : VX020326.D
 Acq On : 23 Dec 2020 10:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 24 00:12:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	50.000	47.842	4.3	112	0.00
3 P	Chloromethane	50.000	44.240	11.5	99	0.00
4 C	Vinyl Chloride	50.000	45.203	9.6#	106	0.00
5 T	Bromomethane	50.000	49.631	0.7	115	0.00
6 T	Chloroethane	50.000	47.083	5.8	103	0.00
7 T	Trichlorofluoromethane	50.000	48.289	3.4	116	0.00
8 T	Diethyl Ether	50.000	47.218	5.6	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.040	3.9	116	0.00
10 T	Methyl Iodide	50.000	42.820	14.4	95	0.00
11 T	Tert butyl alcohol	250.000	232.259	7.1	106	0.00
12 CM	1,1-Dichloroethene	50.000	46.354	7.3#	109	0.00
13 T	Acrolein	250.000	371.685	-48.7#	162	0.00
14 T	Allyl chloride	50.000	46.617	6.8	101	0.00
15 T	Acrylonitrile	250.000	247.162	1.1	102	0.00
16 T	Acetone	250.000	252.861	-1.1	110	0.00
17 T	Carbon Disulfide	50.000	44.282	11.4	104	0.00
18 T	Methyl Acetate	50.000	50.587	-1.2	106	0.00
19 T	Methyl tert-butyl Ether	50.000	48.789	2.4	98	0.00
20 T	Methylene Chloride	50.000	45.095	9.8	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.804	8.4	102	0.00
22 T	Diisopropyl ether	50.000	48.843	2.3	99	0.00
23 T	Vinyl Acetate	250.000	250.710	-0.3	98	0.00
24 P	1,1-Dichloroethane	50.000	47.454	5.1	102	0.00
25 T	2-Butanone	250.000	253.451	-1.4	105	0.00
26 T	2,2-Dichloropropane	50.000	46.810	6.4	104	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.933	6.1	99	0.00
28 T	Bromochloromethane	50.000	46.581	6.8	99	0.00
29 T	Tetrahydrofuran	250.000	249.311	0.3	102	0.00
30 C	Chloroform	50.000	48.125	3.8#	101	0.00
31 T	Cyclohexane	50.000	48.250	3.5	114	0.00
32 T	1,1,1-Trichloroethane	50.000	48.456	3.1	109	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.377	3.2	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	50.707	-1.4	104	0.00
36 T	1,1-Dichloropropene	50.000	48.162	3.7	109	0.00
37 T	Ethyl Acetate	50.000	50.439	-0.9	101	0.00
38 T	Carbon Tetrachloride	50.000	49.385	1.2	110	0.00
39 T	Methylcyclohexane	50.000	50.351	-0.7	118	0.00
40 TM	Benzene	50.000	48.009	4.0	100	0.00
41 T	Methacrylonitrile	50.000	51.870	-3.7	100	0.00
42 TM	1,2-Dichloroethane	50.000	49.674	0.7	99	0.00
43 T	Isopropyl Acetate	50.000	50.678	-1.4	100	0.00
44 TM	Trichloroethene	50.000	47.036	5.9	103	0.00
45 C	1,2-Dichloropropane	50.000	49.538	0.9#	100	0.00
46 T	Dibromomethane	50.000	49.013	2.0	99	0.00
47 T	Bromodichloromethane	50.000	51.404	-2.8	102	0.00
48 T	Methyl methacrylate	50.000	52.053	-4.1	101	0.00
49 T	1,4-Dioxane	1000.000	1046.232	-4.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122320\
 Data File : VX020326.D
 Acq On : 23 Dec 2020 10:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 24 00:12:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.324	1.4	103	0.00
51 T	4-Methyl-2-Pentanone	250.000	264.377	-5.8	103	0.00
52 CM	Toluene	50.000	49.945	0.1#	103	0.00
53 T	t-1,3-Dichloropropene	50.000	51.387	-2.8	100	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.940	-1.9	99	0.00
55 T	1,1,2-Trichloroethane	50.000	50.248	-0.5	100	0.00
56 T	Ethyl methacrylate	50.000	52.811	-5.6	100	0.00
57 T	1,3-Dichloropropane	50.000	50.027	-0.1	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	235.978	5.6	93	0.00
59 T	2-Hexanone	250.000	270.116	-8.0	106	0.00
60 T	Dibromochloromethane	50.000	52.226	-4.5	102	0.00
61 T	1,2-Dibromoethane	50.000	50.785	-1.6	100	0.00
62 S	4-Bromofluorobenzene	50.000	51.017	-2.0	107	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	101	0.00
64 T	Tetrachloroethene	50.000	47.887	4.2	109	0.00
65 PM	Chlorobenzene	50.000	48.136	3.7	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.169	-0.3	102	0.00
67 C	Ethyl Benzene	50.000	50.023	-0.0#	104	0.00
68 T	m/p-Xylenes	100.000	100.615	-0.6	103	0.00
69 T	o-Xylene	50.000	50.000	0.0	101	0.00
70 T	Styrene	50.000	51.523	-3.0	101	0.00
71 P	Bromoform	50.000	53.026	-6.1	107	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	107	0.00
73 T	Isopropylbenzene	50.000	48.238	3.5	107	0.00
74 T	N-aryl acetate	50.000	50.214	-0.4	102	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.092	3.8	104	0.00
76 T	1,2,3-Trichloropropane	50.000	48.767	2.5	106	0.00
77 T	Bromobenzene	50.000	46.332	7.3	104	0.00
78 T	n-propylbenzene	50.000	48.660	2.7	108	0.00
79 T	2-Chlorotoluene	50.000	47.654	4.7	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.586	2.8	105	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.157	-0.3	103	0.00
82 T	4-Chlorotoluene	50.000	47.917	4.2	104	0.00
83 T	tert-Butylbenzene	50.000	49.581	0.8	111	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.953	0.1	105	0.00
85 T	sec-Butylbenzene	50.000	50.429	-0.9	116	0.00
86 T	p-Isopropyltoluene	50.000	50.687	-1.4	112	0.00
87 T	1,3-Dichlorobenzene	50.000	47.577	4.8	104	0.00
88 T	1,4-Dichlorobenzene	50.000	45.645	8.7	103	0.00
89 T	n-Butylbenzene	50.000	51.727	-3.5	117	0.00
90 T	Hexachloroethane	50.000	49.703	0.6	114	0.00
91 T	1,2-Dichlorobenzene	50.000	46.708	6.6	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.687	2.6	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.760	2.5	108	0.00
94 T	Hexachlorobutadiene	50.000	51.666	-3.3	133	0.00
95 T	Naphthalene	50.000	51.273	-2.5	105	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.293	3.4	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122320\
Data File : VX020326.D
Acq On : 23 Dec 2020 10:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 24 00:12:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6