

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX122320\
 Data File : VX020351.D
 Acq On : 23 Dec 2020 20:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 24 00:21:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.01
2 T	Dichlorodifluoromethane	50.000	49.035	1.9	106	0.00
3 P	Chloromethane	50.000	49.536	0.9	102	0.00
4 C	Vinyl Chloride	50.000	48.877	2.2#	106	0.00
5 T	Bromomethane	50.000	50.961	-1.9	109	0.00
6 T	Chloroethane	50.000	50.681	-1.4	102	0.00
7 T	Trichlorofluoromethane	50.000	51.097	-2.2	113	0.00
8 T	Diethyl Ether	50.000	49.591	0.8	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.449	-0.9	112	0.00
10 T	Methyl Iodide	50.000	48.612	2.8	100	0.00
11 T	Tert butyl alcohol	250.000	246.352	1.5	104	0.00
12 CM	1,1-Dichloroethene	50.000	50.074	-0.1#	109	0.00
13 T	Acrolein	250.000	316.498	-26.6#	127	0.00
14 T	Allyl chloride	50.000	48.552	2.9	97	0.00
15 T	Acrylonitrile	250.000	261.589	-4.6	99	0.00
16 T	Acetone	250.000	234.117	6.4	94	0.00
17 T	Carbon Disulfide	50.000	47.112	5.8	103	0.00
18 T	Methyl Acetate	50.000	54.249	-8.5	104	0.00
19 T	Methyl tert-butyl Ether	50.000	50.303	-0.6	94	0.00
20 T	Methylene Chloride	50.000	47.818	4.4	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.186	3.6	99	0.00
22 T	Diisopropyl ether	50.000	51.506	-3.0	97	0.00
23 T	Vinyl Acetate	250.000	259.432	-3.8	94	0.00
24 P	1,1-Dichloroethane	50.000	50.565	-1.1	100	0.00
25 T	2-Butanone	250.000	259.705	-3.9	99	0.01
26 T	2,2-Dichloropropane	50.000	44.069	11.9	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.994	2.0	96	0.00
28 T	Bromochloromethane	50.000	50.541	-1.1	99	0.00
29 T	Tetrahydrofuran	250.000	266.736	-6.7	100	0.00
30 C	Chloroform	50.000	51.371	-2.7#	100	0.00
31 T	Cyclohexane	50.000	49.914	0.2	109	0.01
32 T	1,1,1-Trichloroethane	50.000	51.762	-3.5	107	0.00
33 S	1,2-Dichloroethane-d4	50.000	51.045	-2.1	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	50.972	-1.9	100	0.00
36 T	1,1-Dichloropropene	50.000	49.748	0.5	107	0.00
37 T	Ethyl Acetate	50.000	51.586	-3.2	98	0.00
38 T	Carbon Tetrachloride	50.000	51.481	-3.0	109	0.00
39 T	Methylcyclohexane	50.000	48.248	3.5	108	0.00
40 TM	Benzene	50.000	49.662	0.7	98	0.00
41 T	Methacrylonitrile	50.000	52.685	-5.4	97	0.00
42 TM	1,2-Dichloroethane	50.000	50.593	-1.2	96	0.00
43 T	Isopropyl Acetate	50.000	50.860	-1.7	96	0.00
44 TM	Trichloroethene	50.000	48.364	3.3	100	0.00
45 C	1,2-Dichloropropane	50.000	51.287	-2.6#	99	0.00
46 T	Dibromomethane	50.000	49.062	1.9	94	0.00
47 T	Bromodichloromethane	50.000	52.227	-4.5	98	0.00
48 T	Methyl methacrylate	50.000	52.625	-5.3	97	0.00
49 T	1,4-Dioxane	1000.000	1089.227	-8.9	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122320\
 Data File : VX020351.D
 Acq On : 23 Dec 2020 20:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 24 00:21:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
50 S	Toluene-d8	50.000	52.033	-4.1	104	0.00
51 T	4-Methyl-2-Pentanone	250.000	278.708	-11.5	103	0.00
52 CM	Toluene	50.000	51.799	-3.6#	101	0.00
53 T	t-1,3-Dichloropropene	50.000	51.112	-2.2	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.884	-1.8	94	0.00
55 T	1,1,2-Trichloroethane	50.000	51.511	-3.0	98	0.00
56 T	Ethyl methacrylate	50.000	54.506	-9.0	99	0.00
57 T	1,3-Dichloropropane	50.000	51.839	-3.7	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	239.767	4.1	90	0.00
59 T	2-Hexanone	250.000	280.267	-12.1	104	0.00
60 T	Dibromochloromethane	50.000	53.353	-6.7	99	0.00
61 T	1,2-Dibromoethane	50.000	51.034	-2.1	96	0.00
62 S	4-Bromofluorobenzene	50.000	51.650	-3.3	103	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	48.479	3.0	107	0.00
65 PM	Chlorobenzene	50.000	48.186	3.6	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.594	-1.2	100	0.00
67 C	Ethyl Benzene	50.000	50.578	-1.2#	103	0.00
68 T	m/p-Xylenes	100.000	101.734	-1.7	101	0.00
69 T	o-Xylene	50.000	50.714	-1.4	99	0.00
70 T	Styrene	50.000	51.930	-3.9	99	0.00
71 P	Bromoform	50.000	50.962	-1.9	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	51.099	-2.2	106	0.00
74 T	N-aryl acetate	50.000	51.740	-3.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.666	-3.3	103	0.00
76 T	1,2,3-Trichloropropane	50.000	50.544	-1.1	102	0.00
77 T	Bromobenzene	50.000	47.646	4.7	99	0.00
78 T	n-propylbenzene	50.000	51.231	-2.5	106	0.00
79 T	2-Chlorotoluene	50.000	49.814	0.4	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.175	-2.3	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.994	2.0	94	0.00
82 T	4-Chlorotoluene	50.000	50.088	-0.2	101	0.00
83 T	tert-Butylbenzene	50.000	50.739	-1.5	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.388	-2.8	100	0.00
85 T	sec-Butylbenzene	50.000	51.680	-3.4	110	0.00
86 T	p-Isopropyltoluene	50.000	51.386	-2.8	105	0.00
87 T	1,3-Dichlorobenzene	50.000	48.323	3.4	98	0.00
88 T	1,4-Dichlorobenzene	50.000	46.355	7.3	97	0.00
89 T	n-Butylbenzene	50.000	50.287	-0.6	106	0.00
90 T	Hexachloroethane	50.000	50.049	-0.1	106	0.00
91 T	1,2-Dichlorobenzene	50.000	48.389	3.2	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.937	-1.9	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.107	5.8	97	0.00
94 T	Hexachlorobutadiene	50.000	49.285	1.4	118	0.00
95 T	Naphthalene	50.000	51.989	-4.0	98	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.168	3.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122320\
Data File : VX020351.D
Acq On : 23 Dec 2020 20:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 24 00:21:54 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6