

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122322\
 Data File : VX033554.D
 Acq On : 23 Dec 2022 10:52
 Operator : JC/MD
 Sample : VX1223MBL01
 Misc : 5.00g/5.0mL/100ul/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK605

Quant Time: Dec 24 00:28:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122222WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 24 00:26:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	219885	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	193559	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	65529	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	50512	43.184	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.360%
7) Chloroethane-d5	1.672	69	39242	42.898	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.800%
11) 1,1-Dichloroethene-d2	2.307	63	93274	49.946	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.900%
21) 2-Butanone-d5	4.459	46	83875	98.716	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.720%
24) Chloroform-d	5.056	84	134794	46.484	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.960%
26) 1,2-Dichloroethane-d4	5.952	65	90563	50.370	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.740%
32) Benzene-d6	5.970	84	258356	46.737	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.480%
36) 1,2-Dichloropropane-d6	7.306	67	78069	48.340	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.680%
41) Toluene-d8	8.647	98	246433	45.932	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.860%
43) trans-1,3-Dichloroprop...	8.952	79	35923	46.011	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.020%
47) 2-Hexanone-d5	9.385	63	63927	96.557	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.560%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	102958	46.493	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.980%
66) 1,2-Dichlorobenzene-d4	12.323	152	66859	49.185	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122322\
Data File : VX033554.D
Acq On : 23 Dec 2022 10:52
Operator : JC/MD
Sample : VX1223MBL01
Misc : 5.00g/5.0mL/100ul/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK605

Quant Time: Dec 24 00:28:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM12222WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 24 00:26:44 2022
Response via : Initial Calibration

