

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
 Data File : VX044472.D
 Acq On : 23 Dec 2024 12:12
 Operator : JC/MD
 Sample : P5317-17 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-06-20241217

Quant Time: Dec 23 21:54:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	152527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	266136	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	230576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	107524	40.206	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	80.420%
35) Dibromofluoromethane	5.385	113	83943	45.541	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.080%
50) Toluene-d8	8.647	98	316568	50.907	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.820%
62) 4-Bromofluorobenzene	11.079	95	102749	47.286	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.580%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
Data File : VX044472.D
Acq On : 23 Dec 2024 12:12
Operator : JC/MD
Sample : P5317-17 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-06-20241217

Quant Time: Dec 23 21:54:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

