

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
 Data File : VX044475.D
 Acq On : 23 Dec 2024 13:21
 Operator : JC/MD
 Sample : P5317-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D2-20241217

Quant Time: Dec 23 21:55:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	154378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	267929	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	228380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100385	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	138432	51.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.280%
35) Dibromofluoromethane	5.385	113	106692	57.496	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	115.000%
50) Toluene-d8	8.647	98	389462	62.210	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	124.420%#
62) 4-Bromofluorobenzene	11.079	95	124939	57.114	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	114.220%
Target Compounds						Qvalue
3) Chloromethane	1.294	50	781	0.342	ug/l	89
16) Acetone	2.386	43	1117	1.012	ug/l #	83
20) Methylene Chloride	2.788	84	524	0.253	ug/l #	83
44) Trichloroethene	7.135	130	502	0.276	ug/l	50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
Data File : VX044475.D
Acq On : 23 Dec 2024 13:21
Operator : JC/MD
Sample : P5317-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D2-20241217

Quant Time: Dec 23 21:55:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

