

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
 Data File : VX044481.D
 Acq On : 23 Dec 2024 15:38
 Operator : JC/MD
 Sample : P5374-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190A-GW-748-750

Quant Time: Dec 23 21:58:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	132243	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	237676	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	212650	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	92633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	117110	50.508	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.020%			
35) Dibromofluoromethane	5.379	113	91062	55.319	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.640%			
50) Toluene-d8	8.647	98	333785	60.103	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 120.200%#			
62) 4-Bromofluorobenzene	11.079	95	110042	56.707	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 113.420%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	19823	12.283	ug/l	95
12) 1,1-Dichloroethene	2.312	96	7751	5.229	ug/l	83
16) Acetone	2.379	43	1316	1.392	ug/l	97
24) 1,1-Dichloroethane	3.605	63	1921	0.620	ug/l #	94
27) cis-1,2-Dichloroethene	4.495	96	2405	1.202	ug/l	88
30) Chloroform	5.092	83	1565	0.449	ug/l	94
38) Carbon Tetrachloride	5.678	117	1647	0.719	ug/l #	82
44) Trichloroethene	7.116	130	216035	134.094	ug/l	99
55) 1,1,2-Trichloroethane	9.159	97	597	0.370	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
Data File : VX044481.D
Acq On : 23 Dec 2024 15:38
Operator : JC/MD
Sample : P5374-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-GW-748-750

Quant Time: Dec 23 21:58:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

