

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
 Data File : VX044486.D
 Acq On : 23 Dec 2024 17:31
 Operator : JC/MD
 Sample : P5317-22 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D7-20241217

Quant Time: Dec 23 22:13:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	146702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	250392	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	212795	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	100422	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	108155	42.048	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.100%
35) Dibromofluoromethane	5.379	113	87011	50.174	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.340%
50) Toluene-d8	8.647	98	310122	53.006	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.020%
62) 4-Bromofluorobenzene	11.079	95	105608	51.658	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.320%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	778	0.435	ug/l #	89
16) Acetone	2.386	43	458	0.437	ug/l #	78
44) Trichloroethene	7.123	130	48195	28.396	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
Data File : VX044486.D
Acq On : 23 Dec 2024 17:31
Operator : JC/MD
Sample : P5317-22 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D7-20241217

Quant Time: Dec 23 22:13:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

