

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
 Data File : VX044488.D
 Acq On : 23 Dec 2024 18:17
 Operator : JC/MD
 Sample : P5317-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-05-20241217

Quant Time: Dec 23 22:14:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	132450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	232737	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	95898	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104226	44.881	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	89.760%		
35) Dibromofluoromethane	5.379	113	79952	49.601	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.200%		
50) Toluene-d8	8.647	98	302195	55.570	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	111.140%		
62) 4-Bromofluorobenzene	11.079	95	100909	53.104	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	106.200%		
Target Compounds						
24) 1,1-Dichloroethane	3.593	63	3842	1.238	ug/l	94
44) Trichloroethene	7.129	130	510	0.323	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122324\
Data File : VX044488.D
Acq On : 23 Dec 2024 18:17
Operator : JC/MD
Sample : P5317-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-05-20241217

Quant Time: Dec 23 22:14:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

