

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122418\
 Data File : VX006785.D
 Acq On : 24 Dec 2018 10:52
 Operator : JC/MD
 Sample : J6532-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 25 02:35:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	633289	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	964996	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	885375	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	428118	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	331881	48.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.26%	
35) Dibromofluoromethane	5.51	113	296259	48.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.10%	
50) Toluene-d8	8.71	98	1136420	49.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.13	95	403507	47.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.78%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122418\
Data File : VX006785.D
Acq On : 24 Dec 2018 10:52
Operator : JC/MD
Sample : J6532-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Dec 25 02:35:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

