

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122418\
 Data File : VX006793.D
 Acq On : 24 Dec 2018 14:24
 Operator : JC/MD
 Sample : J6533-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-1

Quant Time: Dec 26 10:10:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	654973	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	987063	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	904324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	442826	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	338798	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
35) Dibromofluoromethane	5.51	113	292253	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
50) Toluene-d8	8.71	98	1157306	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	412395	47.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.70%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	3.06	59	14632	9.288	ug/l #	88
16) Acetone	2.45	43	58362	17.996	ug/l	99
25) 2-Butanone	4.70	43	67745	14.850	ug/l	98
40) Benzene	6.14	78	71212	2.866	ug/l	94
52) Toluene	8.78	92	34890	2.173	ug/l	100
68) m/p-Xylenes	10.36	106	15220	1.234	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122418\
Data File : VX006793.D
Acq On : 24 Dec 2018 14:24
Operator : JC/MD
Sample : J6533-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-1

Quant Time: Dec 26 10:10:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

