

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122418\
 Data File : VX006794.D
 Acq On : 24 Dec 2018 14:51
 Operator : JC/MD
 Sample : J6192-03
 Misc : 6.86g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HD-01-122018

Quant Time: Dec 26 10:11:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	563264	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	858422	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	788601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	413479	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	284542	46.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.74%	
35) Dibromofluoromethane	5.51	113	238446	43.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.84%	
50) Toluene-d8	8.71	98	998724	48.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.60%	
62) 4-Bromofluorobenzene	11.14	95	368860	49.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.42%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122418\
Data File : VX006794.D
Acq On : 24 Dec 2018 14:51
Operator : JC/MD
Sample : J6192-03
Misc : 6.86g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HD-01-122018

Quant Time: Dec 26 10:11:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

