

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122418\
 Data File : VX006801.D
 Acq On : 24 Dec 2018 17:57
 Operator : JC/MD
 Sample : J6548-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 G-C

Quant Time: Dec 25 04:06:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	735461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1077094	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1009582	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	499358	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	364863	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
35) Dibromofluoromethane	5.51	113	318744	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
50) Toluene-d8	8.71	98	1283727	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
62) 4-Bromofluorobenzene	11.13	95	463002	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
Target Compounds						
16) Acetone	2.45	43	29715	8.160	ug/l	98
18) Methyl Acetate	2.78	43	23959	2.237	ug/l	100
20) Methylene Chloride	2.86	84	37530	4.984	ug/l	92
43) Isopropyl Acetate	6.46	43	35844	2.537	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122418\
Data File : VX006801.D
Acq On : 24 Dec 2018 17:57
Operator : JC/MD
Sample : J6548-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
G-C

Quant Time: Dec 25 04:06:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

