

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122418\
 Data File : VX006805.D
 Acq On : 24 Dec 2018 19:44
 Operator : JC/MD
 Sample : J6533-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-5

Quant Time: Dec 25 04:22:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	714520	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1051292	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	953108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	468978	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	355042	46.10	ug/l	0.00
Spiked Amount			Recovery	=	92.20%	
35) Dibromofluoromethane	5.50	113	305893	46.01	ug/l	0.00
Spiked Amount			Recovery	=	92.02%	
50) Toluene-d8	8.71	98	1235904	49.31	ug/l	0.00
Spiked Amount			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	11.13	95	441170	48.06	ug/l	0.00
Spiked Amount			Recovery	=	96.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.07	59	228833	133.150	ug/l	98
16) Acetone	2.45	43	231518	65.438	ug/l	99
18) Methyl Acetate	2.78	43	111551	10.723	ug/l	100
25) 2-Butanone	4.70	43	2679124	538.343	ug/l	98
40) Benzene	6.14	78	49313	1.863	ug/l	96
49) 1,4-Dioxane	7.75	88	3759	0.621	ug/l	90
52) Toluene	8.79	92	43044	2.517	ug/l	98
67) Ethyl Benzene	10.25	91	12836	0.390	ug/l	98
68) m/p-Xylenes	10.35	106	29893	2.299	ug/l	99
69) o-Xylene	10.69	106	9246	0.720	ug/l	99
70) Styrene	10.71	104	50574	2.490	ug/l	95
84) 1,2,4-Trimethylbenzene	11.80	105	15269	0.549	ug/l	73
95) Naphthalene	13.83	128	20276	0.589	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122418\
Data File : VX006805.D
Acq On : 24 Dec 2018 19:44
Operator : JC/MD
Sample : J6533-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-5

Quant Time: Dec 25 04:22:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

