

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX122418\
 Data File : VX006781.D
 Acq On : 24 Dec 2018 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 02:16:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	114	0.00
2 T	Dichlorodifluoromethane	0.457	0.406	11.2	105	0.00
3 P	Chloromethane	0.493	0.433	12.2	105	0.00
4 C	Vinyl Chloride	0.515	0.462	10.3#	107	0.00
5 T	Bromomethane	0.560	0.456	18.6	106	0.00
6 T	Chloroethane	0.356	0.291	18.3	104	0.00
7 T	Trichlorofluoromethane	0.802	0.742	7.5	108	0.00
8 T	Diethyl Ether	0.319	0.285	10.7	108	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.471	0.447	5.1	112	0.00
10 T	Methyl Iodide	0.674	0.642	4.7	102	0.00
11 T	Tert butyl alcohol	0.120	0.109	9.2	106	0.00
12 CM	1,1-Dichloroethene	0.447	0.401	10.3#	107	0.00
13 T	Acrolein	0.081	0.066	18.5	97	0.00
14 T	Allyl chloride	0.752	0.704	6.4	109	0.00
15 T	Acrylonitrile	0.267	0.247	7.5	107	0.00
16 T	Acetone	0.248	0.267	-7.7	128	0.00
17 T	Carbon Disulfide	1.182	0.998	15.6	102	0.00
18 T	Methyl Acetate	0.728	0.705	3.2	113	0.00
19 T	Methyl tert-butyl Ether	1.483	1.409	5.0	110	0.00
20 T	Methylene Chloride	0.512	0.456	10.9	109	0.00
21 T	trans-1,2-Dichloroethene	0.495	0.438	11.5	109	0.00
22 T	Diisopropyl ether	1.376	1.401	-1.8	117	0.00
23 T	Vinyl Acetate	1.125	1.154	-2.6	118	0.00
24 P	1,1-Dichloroethane	0.821	0.776	5.5	115	0.00
25 T	2-Butanone	0.348	0.350	-0.6	117	0.00
26 T	2,2-Dichloropropane	0.634	0.621	2.1	114	0.00
27 T	cis-1,2-Dichloroethene	0.546	0.512	6.2	109	0.00
28 T	Bromochloromethane	0.378	0.400	-5.8	129	0.00
29 T	Tetrahydrofuran	0.229	0.214	6.6	108	0.00
30 C	Chloroform	0.850	0.810	4.7#	112	0.00
31 T	Cyclohexane	0.746	0.690	7.5	108	0.00
32 T	1,1,1-Trichloroethane	0.718	0.700	2.5	109	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.549	-1.9	116	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.316	0.353	-11.7	119	0.00
36 T	1,1-Dichloropropene	0.414	0.409	1.2	108	0.00
37 T	Ethyl Acetate	0.431	0.424	1.6	110	0.00
38 T	Carbon Tetrachloride	0.402	0.426	-6.0	114	0.00
39 T	Methylcyclohexane	0.545	0.542	0.6	111	0.00
40 TM	Benzene	1.259	1.239	1.6	108	0.00
41 T	Methacrylonitrile	0.229	0.237	-3.5	108	0.00
42 TM	1,2-Dichloroethane	0.429	0.417	2.8	110	0.00
43 T	Isopropyl Acetate	0.656	0.699	-6.6	113	0.00
44 TM	Trichloroethene	0.381	0.370	2.9	110	0.00
45 C	1,2-Dichloropropane	0.306	0.312	-2.0#	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX122418\
 Data File : VX006781.D
 Acq On : 24 Dec 2018 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 02:16:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.225	0.223	0.9	108	0.00
47 T	Bromodichloromethane	0.374	0.405	-8.3	115	0.00
48 T	Methyl methacrylate	0.334	0.362	-8.4	113	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	105	0.00
50 S	Toluene-d8	1.192	1.313	-10.2	118	0.00
51 T	4-Methyl-2-Pentanone	0.417	0.448	-7.4	112	0.00
52 CM	Toluene	0.813	0.816	-0.4#	109	0.00
53 T	t-1,3-Dichloropropene	0.401	0.457	-14.0	115	0.00
54 T	cis-1,3-Dichloropropene	0.440	0.500	-13.6	114	0.00
55 T	1,1,2-Trichloroethane	0.333	0.341	-2.4	112	0.00
56 T	Ethyl methacrylate	0.470	0.508	-8.1	110	0.00
57 T	1,3-Dichloropropane	0.536	0.542	-1.1	111	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.270	-8.0	116	0.00
59 T	2-Hexanone	0.319	0.356	-11.6	116	0.00
60 T	Dibromochloromethane	0.307	0.365	-18.9	119	0.00
61 T	1,2-Dibromoethane	0.354	0.366	-3.4	109	0.00
62 S	4-Bromofluorobenzene	0.437	0.491	-12.4	117	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	108	0.00
64 T	Tetrachloroethene	0.396	0.371	6.3	107	0.00
65 PM	Chlorobenzene	1.058	1.036	2.1	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.330	0.370	-12.1	116	0.00
67 C	Ethyl Benzene	1.729	1.733	-0.2#	111	0.00
68 T	m/p-Xylenes	0.682	0.694	-1.8	111	0.00
69 T	o-Xylene	0.674	0.679	-0.7	109	0.00
70 T	Styrene	1.066	1.114	-4.5	110	0.00
71 P	Bromoform	0.251	0.300	-19.5	121	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.421	3.337	2.5	112	0.00
74 T	N-amyl acetate	1.224	1.309	-6.9	118	0.00
75 P	1,1,2,2-Tetrachloroethane	1.017	1.009	0.8	112	0.00
76 T	1,2,3-Trichloropropane	0.929	0.942	-1.4	110	0.00
77 T	Bromobenzene	0.925	0.904	2.3	111	0.00
78 T	n-propylbenzene	3.767	3.815	-1.3	115	0.00
79 T	2-Chlorotoluene	2.283	2.203	3.5	112	0.00
80 T	1,3,5-Trimethylbenzene	2.817	2.821	-0.1	113	0.00
81 T	trans-1,4-Dichloro-2-butene	0.259	0.293	-13.1	120	0.00
82 T	4-Chlorotoluene	2.618	2.610	0.3	114	0.00
83 T	tert-Butylbenzene	2.820	2.828	-0.3	111	0.00
84 T	1,2,4-Trimethylbenzene	2.963	2.865	3.3	112	0.00
85 T	sec-Butylbenzene	3.457	3.473	-0.5	114	0.00
86 T	p-Isopropyltoluene	3.039	3.131	-3.0	115	0.00
87 T	1,3-Dichlorobenzene	1.667	1.663	0.2	114	0.00
88 T	1,4-Dichlorobenzene	1.709	1.671	2.2	114	0.00
89 T	n-Butylbenzene	2.586	2.758	-6.7	119	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX122418\
Data File : VX006781.D
Acq On : 24 Dec 2018 09:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 25 02:16:26 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.414	0.477	-15.2	123	0.00
91 T	1,2-Dichlorobenzene	1.667	1.623	2.6	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.210	0.211	-0.5	111	0.00
93 T	1,2,4-Trichlorobenzene	1.150	1.204	-4.7	119	0.00
94 T	Hexachlorobutadiene	0.542	0.579	-6.8	126	0.00
95 T	Naphthalene	3.671	3.548	3.4	110	0.00
96 T	1,2,3-Trichlorobenzene	1.187	1.188	-0.1	118	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6