


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122418\
Data File : VX006780.D
Acq On : 24 Dec 2018 08:25
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Title : SW846 8260
Last Update : Wed Dec 19 15:20:58 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1627

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	16188	PASS
75	95	30	60	48.0	46530	PASS
95	95	100	100	100.0	97034	PASS
96	95	5	9	6.6	6437	PASS
173	174	0.00	2	1.1	909	PASS
174	95	50	100	86.9	84296	PASS
175	174	5	9	7.5	6351	PASS
176	174	95	101	95.8	80794	PASS
177	176	5	9	6.2	4987	PASS