

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122419\
 Data File : VX014239.D
 Acq On : 24 Dec 2019 12:55
 Operator : JC/SP
 Sample : K6401-01 40X
 Misc : 4.72g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B37

Quant Time: Dec 30 06:46:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	545738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	846897	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	774279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	359142	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	303026	48.99	ug/l	0.00
Spiked Amount			Recovery	=	97.98%	
35) Dibromofluoromethane	5.49	113	248983	48.36	ug/l	0.00
Spiked Amount			Recovery	=	96.72%	
50) Toluene-d8	8.71	98	1003766	50.08	ug/l	0.00
Spiked Amount			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	11.13	95	359623	49.08	ug/l	0.00
Spiked Amount			Recovery	=	98.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	7.20	130	57961	8.777	ug/l	99
67) Ethyl Benzene	10.25	91	610024	21.553	ug/l	100
68) m/p-Xylenes	10.36	106	838419	77.018	ug/l	100
69) o-Xylene	10.70	106	310072	29.042	ug/l	98
73) Isopropylbenzene	11.01	105	30143	1.192	ug/l	97
78) n-propylbenzene	11.35	91	54325	1.915	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	159951	7.516	ug/l	99
83) tert-Butylbenzene	11.76	119	18055	0.872	ug/l	88
84) 1,2,4-Trimethylbenzene	11.80	105	285716	13.406	ug/l	99
85) sec-Butylbenzene	11.94	105	29012	1.187	ug/l	99
86) p-Isopropyltoluene	12.06	119	29682	1.327	ug/l	96
89) n-Butylbenzene	12.38	91	46866	2.456	ug/l #	31
95) Naphthalene	13.83	128	23076	0.984	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122419\
Data File : VX014239.D
Acq On : 24 Dec 2019 12:55
Operator : JC/SP
Sample : K6401-01 40X
Misc : 4.72g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B37

Quant Time: Dec 30 06:46:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

