

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122419\
 Data File : VX014240.D
 Acq On : 24 Dec 2019 13:19
 Operator : JC/SP
 Sample : K6401-04 20X
 Misc : 4.58g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B38

Quant Time: Dec 25 06:55:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	531557	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	818274	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	749098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	343261	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	299505	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
35) Dibromofluoromethane	5.49	113	242985	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	8.71	98	976045	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.13	95	343680	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
Target Compounds						
44) Trichloroethene	7.21	130	9616	1.507	ug/l	90
52) Toluene	8.78	92	32566	2.230	ug/l	100
67) Ethyl Benzene	10.25	91	987033	36.045	ug/l	100
68) m/p-Xylenes	10.36	106	1279261	121.464	ug/l	100
69) o-Xylene	10.70	106	298562	28.904	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122419\
Data File : VX014240.D
Acq On : 24 Dec 2019 13:19
Operator : JC/SP
Sample : K6401-04 20X
Misc : 4.58g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B38

Quant Time: Dec 25 06:55:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

