

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122419\
 Data File : VX014241.D
 Acq On : 24 Dec 2019 13:42
 Operator : JC/SP
 Sample : K6401-05
 Misc : 4.53g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0B39

Quant Time: Dec 25 06:58:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	524588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	810276	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	725929	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	346510	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	291987	49.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.22%	
35) Dibromofluoromethane	5.50	113	228828	46.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.92%	
50) Toluene-d8	8.71	98	966075	50.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.76%	
62) 4-Bromofluorobenzene	11.14	95	334941	47.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122419\
Data File : VX014241.D
Acq On : 24 Dec 2019 13:42
Operator : JC/SP
Sample : K6401-05
Misc : 4.53g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0B39

Quant Time: Dec 25 06:58:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

