

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122419\
 Data File : VX014245.D
 Acq On : 24 Dec 2019 15:15
 Operator : JC/SP
 Sample : K6406-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C-WATER_12.20.19

Quant Time: Dec 26 14:15:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	538616	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	821453	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	744893	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	347884	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	296714	48.61	ug/l	0.00
Spiked Amount			Recovery	=	97.22%	
35) Dibromofluoromethane	5.49	113	245850	49.23	ug/l	0.00
Spiked Amount			Recovery	=	98.46%	
50) Toluene-d8	8.71	98	975965	50.20	ug/l	0.00
Spiked Amount			Recovery	=	100.40%	
62) 4-Bromofluorobenzene	11.13	95	334869	47.12	ug/l	0.00
Spiked Amount			Recovery	=	94.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	24051	6.069	ug/l	95
17) Carbon Disulfide	2.56	76	12034	1.574	ug/l #	92
25) 2-Butanone	4.67	43	11835	2.422	ug/l #	84
67) Ethyl Benzene	10.25	91	12603	0.463	ug/l #	84
68) m/p-Xylenes	10.35	106	4187	0.400	ug/l	100
70) Styrene	10.71	104	12909	0.743	ug/l	95
87) 1,3-Dichlorobenzene	12.03	146	3289	0.278	ug/l	99
89) n-Butylbenzene	12.39	91	4719	0.255	ug/l	71
91) 1,2-Dichlorobenzene	12.39	146	3223	0.271	ug/l	88
93) 1,2,4-Trichlorobenzene	13.64	180	4577	0.592	ug/l #	91
95) Naphthalene	13.83	128	13515	0.595	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122419\
Data File : VX014245.D
Acq On : 24 Dec 2019 15:15
Operator : JC/SP
Sample : K6406-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C-WATER_12.20.19

Quant Time: Dec 26 14:15:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

