

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122419\
 Data File : VX014248.D
 Acq On : 24 Dec 2019 16:25
 Operator : JC/SP
 Sample : K6405-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 992-MW-02-(23.8)

Quant Time: Dec 25 07:41:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	554748	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	845724	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	758548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	339574	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	301887	48.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.04%	
35) Dibromofluoromethane	5.49	113	249508	48.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.06%	
50) Toluene-d8	8.71	98	1005344	50.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.46%	
62) 4-Bromofluorobenzene	11.13	95	336587	46.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.00%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	12145	1.728	ug/l	92
16) Acetone	2.43	43	14393	3.526	ug/l	94
21) trans-1,2-Dichloroethene	3.16	96	8361	1.422	ug/l	93
27) cis-1,2-Dichloroethene	4.59	96	241676	36.349	ug/l	90
44) Trichloroethene	7.21	130	237229	35.971	ug/l	98
64) Tetrachloroethene	9.33	164	434010	64.660	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122419\
Data File : VX014248.D
Acq On : 24 Dec 2019 16:25
Operator : JC/SP
Sample : K6405-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
992-MW-02-(23.8)

Quant Time: Dec 25 07:41:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

