

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122420\
 Data File : VX020376.D
 Acq On : 24 Dec 2020 10:18
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 01:08:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	112	0.00
2 T	Dichlorodifluoromethane	50.000	48.459	3.1	115	0.00
3 P	Chloromethane	50.000	45.567	8.9	103	0.00
4 C	Vinyl Chloride	50.000	46.331	7.3#	111	0.00
5 T	Bromomethane	50.000	59.656	-19.3	140	0.00
6 T	Chloroethane	50.000	47.729	4.5	106	0.00
7 T	Trichlorofluoromethane	50.000	48.990	2.0	119	0.00
8 T	Diethyl Ether	50.000	46.087	7.8	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.391	3.2	118	0.00
10 T	Methyl Iodide	50.000	45.967	8.1	104	0.00
11 T	Tert butyl alcohol	250.000	228.432	8.6	106	0.00
12 CM	1,1-Dichloroethene	50.000	46.688	6.6#	112	0.00
13 T	Acrolein	250.000	261.332	-4.5	116	0.00
14 T	Allyl chloride	50.000	46.929	6.1	103	0.00
15 T	Acrylonitrile	250.000	244.681	2.1	102	0.00
16 T	Acetone	250.000	275.871	-10.3	122	0.00
17 T	Carbon Disulfide	50.000	44.835	10.3	107	0.00
18 T	Methyl Acetate	50.000	50.549	-1.1	107	0.00
19 T	Methyl tert-butyl Ether	50.000	48.385	3.2	99	0.00
20 T	Methylene Chloride	50.000	44.900	10.2	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.253	7.5	105	0.00
22 T	Diisopropyl ether	50.000	48.738	2.5	100	0.00
23 T	Vinyl Acetate	250.000	247.555	1.0	99	0.00
24 P	1,1-Dichloroethane	50.000	48.000	4.0	104	0.00
25 T	2-Butanone	250.000	260.613	-4.2	109	0.00
26 T	2,2-Dichloropropane	50.000	49.238	1.5	112	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.244	5.5	101	0.00
28 T	Bromochloromethane	50.000	46.501	7.0	100	0.00
29 T	Tetrahydrofuran	250.000	248.086	0.8	103	0.00
30 C	Chloroform	50.000	48.471	3.1#	103	0.00
31 T	Cyclohexane	50.000	49.109	1.8	118	0.00
32 T	1,1,1-Trichloroethane	50.000	49.804	0.4	114	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.154	3.7	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	103	0.00
35 S	Dibromofluoromethane	50.000	50.437	-0.9	104	0.00
36 T	1,1-Dichloropropene	50.000	50.108	-0.2	114	0.00
37 T	Ethyl Acetate	50.000	51.049	-2.1	103	0.00
38 T	Carbon Tetrachloride	50.000	51.830	-3.7	116	0.00
39 T	Methylcyclohexane	50.000	51.822	-3.6	122	0.00
40 TM	Benzene	50.000	49.310	1.4	103	0.00
41 T	Methacrylonitrile	50.000	51.744	-3.5	100	0.00
42 TM	1,2-Dichloroethane	50.000	50.169	-0.3	101	0.00
43 T	Isopropyl Acetate	50.000	50.746	-1.5	101	0.00
44 TM	Trichloroethene	50.000	48.596	2.8	106	0.00
45 C	1,2-Dichloropropane	50.000	50.538	-1.1#	103	0.00
46 T	Dibromomethane	50.000	48.337	3.3	98	0.00
47 T	Bromodichloromethane	50.000	51.893	-3.8	103	0.00
48 T	Methyl methacrylate	50.000	51.934	-3.9	101	0.00
49 T	1,4-Dioxane	1000.000	1023.449	-2.3	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122420\
 Data File : VX020376.D
 Acq On : 24 Dec 2020 10:18
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 01:08:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
50 S	Toluene-d8	50.000	51.006	-2.0	107	0.00
51 T	4-Methyl-2-Pentanone	250.000	273.015	-9.2	107	0.00
52 CM	Toluene	50.000	51.536	-3.1#	106	0.00
53 T	t-1,3-Dichloropropene	50.000	53.138	-6.3	103	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.851	-5.7	103	0.00
55 T	1,1,2-Trichloroethane	50.000	50.215	-0.4	100	0.00
56 T	Ethyl methacrylate	50.000	53.622	-7.2	102	0.00
57 T	1,3-Dichloropropane	50.000	51.451	-2.9	102	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	246.893	1.2	97	0.00
59 T	2-Hexanone	250.000	280.582	-12.2	110	0.00
60 T	Dibromochloromethane	50.000	52.862	-5.7	103	0.00
61 T	1,2-Dibromoethane	50.000	50.790	-1.6	101	0.00
62 S	4-Bromofluorobenzene	50.000	50.277	-0.6	105	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	51.532	-3.1	114	0.00
65 PM	Chlorobenzene	50.000	50.088	-0.2	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.056	-4.1	104	0.00
67 C	Ethyl Benzene	50.000	52.776	-5.6#	108	0.00
68 T	m/p-Xylenes	100.000	104.153	-4.2	104	0.00
69 T	o-Xylene	50.000	52.894	-5.8	104	0.00
70 T	Styrene	50.000	53.979	-8.0	103	0.00
71 P	Bromoform	50.000	52.786	-5.6	104	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	102	0.00
73 T	Isopropylbenzene	50.000	52.253	-4.5	111	0.00
74 T	N-aryl acetate	50.000	51.808	-3.6	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.826	-1.7	104	0.00
76 T	1,2,3-Trichloropropane	50.000	50.083	-0.2	104	0.00
77 T	Bromobenzene	50.000	48.180	3.6	103	0.00
78 T	n-propylbenzene	50.000	52.858	-5.7	112	0.00
79 T	2-Chlorotoluene	50.000	50.740	-1.5	106	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.545	-5.1	109	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.364	-2.7	101	0.00
82 T	4-Chlorotoluene	50.000	50.822	-1.6	105	0.00
83 T	tert-Butylbenzene	50.000	51.532	-3.1	110	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.494	-5.0	106	0.00
85 T	sec-Butylbenzene	50.000	54.092	-8.2	119	0.00
86 T	p-Isopropyltoluene	50.000	53.965	-7.9	114	0.00
87 T	1,3-Dichlorobenzene	50.000	49.406	1.2	103	0.00
88 T	1,4-Dichlorobenzene	50.000	48.063	3.9	103	0.00
89 T	n-Butylbenzene	50.000	55.549	-11.1	120	0.00
90 T	Hexachloroethane	50.000	53.076	-6.2	116	0.00
91 T	1,2-Dichlorobenzene	50.000	49.294	1.4	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.844	-1.7	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.886	0.2	105	0.00
94 T	Hexachlorobutadiene	50.000	55.534	-11.1	137	0.00
95 T	Naphthalene	50.000	52.394	-4.8	102	0.00
96 T	1,2,3-Trichlorobenzene	50.000	49.656	0.7	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122420\
Data File : VX020376.D
Acq On : 24 Dec 2020 10:18
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 25 01:08:28 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6