

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX122420\
 Data File : VX020399.D
 Acq On : 24 Dec 2020 19:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 01:17:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	103	0.01
2 T	Dichlorodifluoromethane	50.000	48.176	3.6	106	0.00
3 P	Chloromethane	50.000	46.366	7.3	97	0.00
4 C	Vinyl Chloride	50.000	47.698	4.6#	105	0.00
5 T	Bromomethane	50.000	52.937	-5.9	115	0.00
6 T	Chloroethane	50.000	48.459	3.1	99	0.00
7 T	Trichlorofluoromethane	50.000	49.579	0.8	111	0.00
8 T	Diethyl Ether	50.000	47.530	4.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.740	2.5	110	0.00
10 T	Methyl Iodide	50.000	48.860	2.3	102	0.00
11 T	Tert butyl alcohol	250.000	259.211	-3.7	110	0.00
12 CM	1,1-Dichloroethene	50.000	48.321	3.4#	106	0.00
13 T	Acrolein	250.000	252.234	-0.9	103	0.00
14 T	Allyl chloride	50.000	48.089	3.8	97	0.00
15 T	Acrylonitrile	250.000	264.906	-6.0	102	0.00
16 T	Acetone	250.000	237.276	5.1	97	0.00
17 T	Carbon Disulfide	50.000	45.043	9.9	99	0.00
18 T	Methyl Acetate	50.000	55.016	-10.0	107	0.00
19 T	Methyl tert-butyl Ether	50.000	49.653	0.7	93	0.00
20 T	Methylene Chloride	50.000	46.480	7.0	94	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.838	6.3	98	0.00
22 T	Diisopropyl ether	50.000	50.810	-1.6	96	0.00
23 T	Vinyl Acetate	250.000	256.901	-2.8	94	0.00
24 P	1,1-Dichloroethane	50.000	49.630	0.7	99	0.00
25 T	2-Butanone	250.000	264.451	-5.8	102	0.01
26 T	2,2-Dichloropropane	50.000	43.825	12.3	91	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.972	4.1	95	0.00
28 T	Bromochloromethane	50.000	49.076	1.8	97	0.00
29 T	Tetrahydrofuran	250.000	273.056	-9.2	104	0.01
30 C	Chloroform	50.000	50.104	-0.2#	98	0.00
31 T	Cyclohexane	50.000	48.335	3.3	107	0.00
32 T	1,1,1-Trichloroethane	50.000	50.886	-1.8	107	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.110	-0.2	98	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	99	0.00
35 S	Dibromofluoromethane	50.000	50.888	-1.8	100	0.00
36 T	1,1-Dichloropropene	50.000	48.393	3.2	105	0.00
37 T	Ethyl Acetate	50.000	52.891	-5.8	101	0.00
38 T	Carbon Tetrachloride	50.000	50.767	-1.5	108	0.00
39 T	Methylcyclohexane	50.000	47.578	4.8	107	0.00
40 TM	Benzene	50.000	48.917	2.2	97	0.00
41 T	Methacrylonitrile	50.000	53.735	-7.5	99	0.00
42 TM	1,2-Dichloroethane	50.000	50.030	-0.1	96	0.00
43 T	Isopropyl Acetate	50.000	51.594	-3.2	98	0.01
44 TM	Trichloroethene	50.000	47.253	5.5	99	0.00
45 C	1,2-Dichloropropane	50.000	50.199	-0.4#	97	0.00
46 T	Dibromomethane	50.000	48.904	2.2	94	0.00
47 T	Bromodichloromethane	50.000	51.256	-2.5	97	0.00
48 T	Methyl methacrylate	50.000	53.310	-6.6	99	0.00
49 T	1,4-Dioxane	1000.000	1108.322	-10.8	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122420\
 Data File : VX020399.D
 Acq On : 24 Dec 2020 19:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 01:17:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
50 S	Toluene-d8	50.000	49.924	0.2	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	276.499	-10.6	103	0.00
52 CM	Toluene	50.000	49.955	0.1#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	49.532	0.9	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.023	-0.0	93	0.00
55 T	1,1,2-Trichloroethane	50.000	50.702	-1.4	97	0.00
56 T	Ethyl methacrylate	50.000	53.020	-6.0	96	0.00
57 T	1,3-Dichloropropane	50.000	50.352	-0.7	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	241.884	3.2	91	0.00
59 T	2-Hexanone	250.000	279.698	-11.9	104	0.00
60 T	Dibromochloromethane	50.000	52.190	-4.4	97	0.00
61 T	1,2-Dibromoethane	50.000	50.412	-0.8	95	0.00
62 S	4-Bromofluorobenzene	50.000	49.798	0.4	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	48.044	3.9	104	0.00
65 PM	Chlorobenzene	50.000	47.837	4.3	95	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.171	-2.3	99	0.00
67 C	Ethyl Benzene	50.000	50.110	-0.2#	100	0.00
68 T	m/p-Xylenes	100.000	100.666	-0.7	98	0.00
69 T	o-Xylene	50.000	50.435	-0.9	97	0.00
70 T	Styrene	50.000	51.286	-2.6	96	0.00
71 P	Bromoform	50.000	52.599	-5.2	101	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	49.424	1.2	103	0.00
74 T	N-aryl acetate	50.000	51.260	-2.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.234	-0.5	101	0.00
76 T	1,2,3-Trichloropropane	50.000	49.143	1.7	100	0.00
77 T	Bromobenzene	50.000	46.671	6.7	98	0.00
78 T	n-propylbenzene	50.000	49.052	1.9	102	0.00
79 T	2-Chlorotoluene	50.000	48.005	4.0	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.401	1.2	100	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.427	3.1	94	0.00
82 T	4-Chlorotoluene	50.000	48.382	3.2	98	0.00
83 T	tert-Butylbenzene	50.000	49.594	0.8	104	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.552	0.9	98	0.00
85 T	sec-Butylbenzene	50.000	49.671	0.7	107	0.00
86 T	p-Isopropyltoluene	50.000	49.037	1.9	102	0.00
87 T	1,3-Dichlorobenzene	50.000	47.193	5.6	97	0.00
88 T	1,4-Dichlorobenzene	50.000	45.757	8.5	96	0.00
89 T	n-Butylbenzene	50.000	49.555	0.9	105	0.00
90 T	Hexachloroethane	50.000	49.980	0.0	107	0.00
91 T	1,2-Dichlorobenzene	50.000	48.048	3.9	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.003	-4.0	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.011	6.0	97	0.00
94 T	Hexachlorobutadiene	50.000	49.834	0.3	121	0.00
95 T	Naphthalene	50.000	51.602	-3.2	99	0.00
96 T	1,2,3-Trichlorobenzene	50.000	49.134	1.7	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122420\
Data File : VX020399.D
Acq On : 24 Dec 2020 19:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 25 01:17:57 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6