

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122421\
 Data File : VX026010.D
 Acq On : 24 Dec 2021 12:16
 Operator : JC/MD
 Sample : VX1224MBL01
 Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK659

Quant Time: Dec 24 16:39:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 23 21:00:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	249036	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	246651	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	119206	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	89886	50.898	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.800%
7) Chloroethane-d5	1.666	69	69985	49.588	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.180%
11) 1,1-Dichloroethene-d2	2.307	63	133553	38.162	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.320%
21) 2-Butanone-d5	4.446	46	139086	90.840	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.840%
24) Chloroform-d	5.056	84	150837	42.680	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.360%
26) 1,2-Dichloroethane-d4	5.952	65	117057	48.920	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.840%
32) Benzene-d6	5.971	84	351422	49.500	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	7.306	67	110212	49.941	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.880%
41) Toluene-d8	8.647	98	331728	50.045	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.080%
43) trans-1,3-Dichloroprop...	8.952	79	45776	47.677	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.360%
47) 2-Hexanone-d5	9.385	63	111116	97.560	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.560%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	150704	46.313	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.620%
66) 1,2-Dichlorobenzene-d4	12.323	152	120474	53.482	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122421\
Data File : VX026010.D
Acq On : 24 Dec 2021 12:16
Operator : JC/MD
Sample : VX1224MBL01
Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK659

Quant Time: Dec 24 16:39:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 23 21:00:32 2021
Response via : Initial Calibration

