

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122521\
 Data File : VX026045.D
 Acq On : 25 Dec 2021 19:38
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.00G/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK619

Quant Time: Dec 27 05:17:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 23 21:00:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	207405	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	189200	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	90834	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	80534	54.756	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	109.520%
7) Chloroethane-d5	1.666	69	68579	58.345	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	116.700%
11) 1,1-Dichloroethene-d2	2.306	63	118167	40.543	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.080%
21) 2-Butanone-d5	4.458	46	153380	120.283	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.280%
24) Chloroform-d	5.062	84	139089	47.256	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.520%
26) 1,2-Dichloroethane-d4	5.964	65	121184	60.811	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	121.620%
32) Benzene-d6	5.976	84	328112	60.250	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	120.500%
36) 1,2-Dichloropropane-d6	7.312	67	106055	62.650	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	125.300%#
41) Toluene-d8	8.653	98	277406	54.557	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.120%
43) trans-1,3-Dichloroprop...	8.952	79	38201	51.869	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.740%
47) 2-Hexanone-d5	9.384	63	119430	136.701	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	136.700%#
56) 1,1,2,2-Tetrachloroeth...	11.195	84	145175	58.161	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	116.320%
66) 1,2-Dichlorobenzene-d4	12.323	152	99482	57.957	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.920%
Target Compounds					Qvalue	
25) Chloroform	5.099	83	29036	9.942	ug/L	99
52) Ethylbenzene	10.195	91	8139	1.173	ug/L	90
53) m,p-Xylene	10.305	106	11511	4.346	ug/L	98
63) 1,2,4-Trimethylbenzene	11.756	105	10655	2.035	ug/L #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122521\
Data File : VX026045.D
Acq On : 25 Dec 2021 19:38
Operator : JC/MD
Sample : VIBLK
Misc : 5.00G/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK619

Quant Time: Dec 27 05:17:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 23 21:00:32 2021
Response via : Initial Calibration

