

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122618\
 Data File : VX006819.D
 Acq On : 26 Dec 2018 15:36
 Operator : JC/MD
 Sample : J6549-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8

Quant Time: Jan 14 13:29:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	800244	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1159437	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1056639	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	553942	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	381562	44.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.48%	
35) Dibromofluoromethane	5.50	113	339609	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
50) Toluene-d8	8.71	98	1355370	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
62) 4-Bromofluorobenzene	11.13	95	500402	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
Target Compounds						
18) Methyl Acetate	2.78	43	22712	1.949	ug/l	96
20) Methylene Chloride	2.86	84	7613	0.929	ug/l #	83
64) Tetrachloroethene	9.34	164	7232	0.864	ug/l	87
84) 1,2,4-Trimethylbenzene	11.80	105	12969	0.395	ug/l	96
95) Naphthalene	13.83	128	9558	0.235	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122618\
Data File : VX006819.D
Acq On : 26 Dec 2018 15:36
Operator : JC/MD
Sample : J6549-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Jan 14 13:29:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

