

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122619\
 Data File : VX014276.D
 Acq On : 26 Dec 2019 17:54
 Operator : JC/SP
 Sample : K6405-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1000-MW-19-(28)

Quant Time: Dec 27 07:14:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	532412	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	820686	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	734410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	317429	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	300179	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
35) Dibromofluoromethane	5.49	113	245448	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
50) Toluene-d8	8.71	98	984854	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
62) 4-Bromofluorobenzene	11.14	95	322687	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
Target Compounds						
16) Acetone	2.43	43	18283	4.667	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	3159	0.495	ug/l	69
44) Trichloroethene	7.22	130	4429	0.692	ug/l	76
64) Tetrachloroethene	9.33	164	7060	1.086	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122619\
Data File : VX014276.D
Acq On : 26 Dec 2019 17:54
Operator : JC/SP
Sample : K6405-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1000-MW-19-(28)

Quant Time: Dec 27 07:14:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

