

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122619\
 Data File : VX014277.D
 Acq On : 26 Dec 2019 18:17
 Operator : JC/SP
 Sample : K6420-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Dec 27 07:17:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	533983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	822739	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	742973	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	328157	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	299075	49.42	ug/l	0.00
Spiked Amount			Recovery	=	98.84%	
35) Dibromofluoromethane	5.49	113	246176	49.22	ug/l	0.00
Spiked Amount			Recovery	=	98.44%	
50) Toluene-d8	8.71	98	980885	50.38	ug/l	0.00
Spiked Amount			Recovery	=	100.76%	
62) 4-Bromofluorobenzene	11.13	95	328533	46.16	ug/l	0.00
Spiked Amount			Recovery	=	92.32%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.38	101	52916	10.483	ug/l	100
16) Acetone	2.43	43	29506	7.510	ug/l	100
44) Trichloroethene	7.21	130	90711	14.139	ug/l	96
64) Tetrachloroethene	9.33	164	60698	9.232	ug/l	96
67) Ethyl Benzene	10.25	91	10265	0.378	ug/l	94
68) m/p-Xylenes	10.35	106	7839	0.750	ug/l	89
69) o-Xylene	10.69	106	4378	0.427	ug/l	95
70) Styrene	10.71	104	17403	1.005	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122619\
Data File : VX014277.D
Acq On : 26 Dec 2019 18:17
Operator : JC/SP
Sample : K6420-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Dec 27 07:17:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

