


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX122619\
Data File : VX014287.D
Acq On : 26 Dec 2019 22:33
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Title : SW846 8260
Last Update : Tue Dec 17 03:01:07 2019



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1672

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	21035	PASS
75	95	30	60	47.5	54592	PASS
95	95	100	100	100.0	114845	PASS
96	95	5	9	6.5	7502	PASS
173	174	0.00	2	0.9	925	PASS
174	95	50	100	91.1	104614	PASS
175	174	5	9	6.9	7247	PASS
176	174	95	101	96.4	100797	PASS
177	176	5	9	6.4	6419	PASS