

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122624\
 Data File : VX044502.D
 Acq On : 26 Dec 2024 12:15
 Operator : JC/MD
 Sample : P5322-02ME
 Misc : 5.36g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YE5X6ME

Quant Time: Dec 27 03:40:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 27 03:37:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	218545	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	195562	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	97464	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	82572	41.287	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	82.580%		
7) Chloroethane-d5	1.648	69	53097	34.273	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	68.540%#		
11) 1,1-Dichloroethene-d2	2.288	65	33606	39.291	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	78.580%		
21) 2-Butanone-d5	4.471	46	82512	67.386	ug/L	0.01
Spiked Amount 100.000	Range 40 - 130		Recovery =	67.390%		
24) Chloroform-d	5.050	84	138283	36.416	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	72.840%		
26) 1,2-Dichloroethane-d4	5.952	65	98056	36.515	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	73.020%		
32) Benzene-d6	5.964	84	287087	42.486	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.980%		
36) 1,2-Dichloropropane-d6	7.305	67	84953	40.489	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	80.980%		
41) Toluene-d8	8.647	98	266893	42.717	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	85.440%		
43) trans-1,3-Dichloroprop...	8.951	79	33471	33.195	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	66.400%		
47) 2-Hexanone-d5	9.390	63	54552	66.681	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	66.680%		
56) 1,1,2,2-Tetrachloroeth...	11.189	84	102949	36.592	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	73.180%		
66) 1,2-Dichlorobenzene-d4	12.317	152	100503	44.258	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.520%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122624\
Data File : VX044502.D
Acq On : 26 Dec 2024 12:15
Operator : JC/MD
Sample : P5322-02ME
Misc : 5.36g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YE5X6ME

Quant Time: Dec 27 03:40:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120524WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 27 03:37:01 2024
Response via : Initial Calibration

