

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122719\
 Data File : VX014307.D
 Acq On : 27 Dec 2019 14:35
 Operator : JC/SP
 Sample : K6457-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 30 06:01:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	526735	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	817568	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	723118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	312236	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	294192	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
35) Dibromofluoromethane	5.49	113	244892	49.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.54%	
50) Toluene-d8	8.71	98	970106	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
62) 4-Bromofluorobenzene	11.13	95	317934	44.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122719\
Data File : VX014307.D
Acq On : 27 Dec 2019 14:35
Operator : JC/SP
Sample : K6457-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Dec 30 06:01:51 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

