

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122719\
 Data File : VX014309.D
 Acq On : 27 Dec 2019 15:21
 Operator : JC/SP
 Sample : K6434-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 30 06:05:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	536044	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	819776	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	743842	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	315360	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	298678	49.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.32%	
35) Dibromofluoromethane	5.49	113	244402	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
50) Toluene-d8	8.71	98	974066	50.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.42%	
62) 4-Bromofluorobenzene	11.14	95	325929	45.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.92%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122719\
Data File : VX014309.D
Acq On : 27 Dec 2019 15:21
Operator : JC/SP
Sample : K6434-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 30 06:05:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

