

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122719\
 Data File : VX014312.D
 Acq On : 27 Dec 2019 16:31
 Operator : JC/SP
 Sample : K6434-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-3

Quant Time: Dec 30 06:12:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	504378	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	784000	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	700898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	299480	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	287260	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
35) Dibromofluoromethane	5.49	113	237379	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.71	98	930515	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
62) 4-Bromofluorobenzene	11.14	95	305444	45.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.06%	
Target Compounds						
16) Acetone	2.43	43	8860	2.387	ug/l #	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122719\
Data File : VX014312.D
Acq On : 27 Dec 2019 16:31
Operator : JC/SP
Sample : K6434-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-3

Quant Time: Dec 30 06:12:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

