

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122719\
 Data File : VX014317.D
 Acq On : 27 Dec 2019 18:28
 Operator : JC/SP
 Sample : K6449-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LW-04

Quant Time: Dec 30 06:22:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	503387	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	777952	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	694420	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	299085	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	279702	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
35) Dibromofluoromethane	5.49	113	237307	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
50) Toluene-d8	8.71	98	928186	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	11.14	95	303805	45.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.28%	
Target Compounds						
16) Acetone	2.43	43	17021	4.595	ug/l	97
27) cis-1,2-Dichloroethene	4.59	96	3175	0.526	ug/l	86
44) Trichloroethene	7.21	130	6155	1.015	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122719\
Data File : VX014317.D
Acq On : 27 Dec 2019 18:28
Operator : JC/SP
Sample : K6449-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LW-04

Quant Time: Dec 30 06:22:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

