

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122719\
 Data File : VX014325.D
 Acq On : 27 Dec 2019 21:34
 Operator : JC/SP
 Sample : K6449-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUPLICATE

Quant Time: Dec 30 06:44:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	490536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	760301	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	672228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	293183	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	274315	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
35) Dibromofluoromethane	5.49	113	226550	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
50) Toluene-d8	8.71	98	901985	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
62) 4-Bromofluorobenzene	11.13	95	297776	45.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.54%	
Target Compounds						
16) Acetone	2.44	43	10956	3.035	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122719\
Data File : VX014325.D
Acq On : 27 Dec 2019 21:34
Operator : JC/SP
Sample : K6449-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUPLICATE

Quant Time: Dec 30 06:44:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

