

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122719\
 Data File : VX014339.D
 Acq On : 28 Dec 2019 04:10
 Operator : JC/SP
 Sample : K6441-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 K321-WC-02

Quant Time: Dec 30 07:07:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	524802	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	819359	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	743224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	327668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	293067	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
35) Dibromofluoromethane	5.49	113	241078	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
50) Toluene-d8	8.71	98	990807	51.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.18%	
62) 4-Bromofluorobenzene	11.13	95	334122	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
Target Compounds						
16) Acetone	2.44	43	22209	5.751	ug/l	100
20) Methylene Chloride	2.85	84	15962	2.625	ug/l #	87
43) Isopropyl Acetate	6.43	43	120672	8.783	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122719\
Data File : VX014339.D
Acq On : 28 Dec 2019 04:10
Operator : JC/SP
Sample : K6441-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
K321-WC-02

Quant Time: Dec 30 07:07:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

