

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
 Data File : VX026059.D
 Acq On : 27 Dec 2021 14:13
 Operator : JC/MD
 Sample : PB141697TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VLEB697

Quant Time: Dec 28 01:30:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 23 21:00:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	197491	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	182939	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	86948	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	89575	63.960	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 127.920%			
7) Chloroethane-d5	1.673	69	66004	58.973	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 117.940%			
11) 1,1-Dichloroethene-d2	2.313	63	124246	44.769	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 89.540%			
21) 2-Butanone-d5	4.465	46	134439	110.722	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 110.720%			
24) Chloroform-d	5.068	84	155500	55.483	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 110.960%			
26) 1,2-Dichloroethane-d4	5.958	65	103123	54.346	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 108.700%			
32) Benzene-d6	5.983	84	318515	60.490	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 120.980%			
36) 1,2-Dichloropropane-d6	7.312	67	97679	59.677	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 119.360%			
41) Toluene-d8	8.653	98	283440	57.652	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 115.300%			
43) trans-1,3-Dichloroprop...	8.952	79	37752	53.013	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 106.020%			
47) 2-Hexanone-d5	9.385	63	93775	111.010	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 111.010%			
56) 1,1,2,2-Tetrachloroeth...	11.195	84	125915	52.171	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 104.340%			
66) 1,2-Dichlorobenzene-d4	12.323	152	97823	59.538	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 119.080%			
Target Compounds					Qvalue	
13) Acetone	2.386	43	1593	2.400	ug/L	67
16) Methylene chloride	2.794	84	6133	3.975	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
Data File : VX026059.D
Acq On : 27 Dec 2021 14:13
Operator : JC/MD
Sample : PB141697TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VLEB697

Quant Time: Dec 28 01:30:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 23 21:00:32 2021
Response via : Initial Calibration

