

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
 Data File : VX026049.D
 Acq On : 27 Dec 2021 10:07
 Operator : JC/MD
 Sample : VX1227MBL01
 Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK660

Quant Time: Dec 27 15:24:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 23 21:00:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	224034	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	209083	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	101370	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	98986	62.306	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 124.620%	
7) Chloroethane-d5	1.672	69	74591	58.750	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 117.500%	
11) 1,1-Dichloroethene-d2	2.306	63	139468	44.300	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 88.600%	
21) 2-Butanone-d5	4.452	46	137054	99.502	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 99.500%	
24) Chloroform-d	5.062	84	130621	41.085	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 82.160%	
26) 1,2-Dichloroethane-d4	5.958	65	112327	52.183	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 104.360%	
32) Benzene-d6	5.970	84	345359	57.386	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 114.780%	
36) 1,2-Dichloropropane-d6	7.305	67	107221	57.316	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 114.640%	
41) Toluene-d8	8.647	98	313465	55.786	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 111.580%	
43) trans-1,3-Dichloroprop...	8.951	79	46003	56.522	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 113.040%	
47) 2-Hexanone-d5	9.384	63	105747	109.529	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 109.530%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	144115	52.246	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 104.500%	
66) 1,2-Dichlorobenzene-d4	12.323	152	110915	57.902	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 115.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
Data File : VX026049.D
Acq On : 27 Dec 2021 10:07
Operator : JC/MD
Sample : VX1227MBL01
Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK660

Quant Time: Dec 27 15:24:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 23 21:00:32 2021
Response via : Initial Calibration

