

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
 Data File : VX026058.D
 Acq On : 27 Dec 2021 13:50
 Operator : JC/MD
 Sample : M5213-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK002

Quant Time: Dec 28 01:26:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 23 21:00:32 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	202894	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	194220	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	91104	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.374	65	94478	65.665	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 131.320%	
7) Chloroethane-d5	1.672	69	72647	63.180	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 126.360%	
11) 1,1-Dichloroethene-d2	2.312	63	130026	45.604	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 91.200%	
21) 2-Butanone-d5	4.458	46	128789	103.244	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 103.240%	
24) Chloroform-d	5.068	84	159771	55.489	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 110.980%	
26) 1,2-Dichloroethane-d4	5.958	65	108367	55.588	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 111.180%	
32) Benzene-d6	5.983	84	325615	58.246	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 116.500%	
36) 1,2-Dichloropropane-d6	7.318	67	104521	60.148	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 120.300%#	
41) Toluene-d8	8.653	98	303081	58.066	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 116.140%	
43) trans-1,3-Dichloroprop...	8.952	79	41412	54.775	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 109.560%	
47) 2-Hexanone-d5	9.384	63	97636	108.867	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 108.870%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	130104	50.776	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 101.560%	
66) 1,2-Dichlorobenzene-d4	12.323	152	105764	61.435	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 122.860%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122721\
Data File : VX026058.D
Acq On : 27 Dec 2021 13:50
Operator : JC/MD
Sample : M5213-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK002

Quant Time: Dec 28 01:26:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 23 21:00:32 2021
Response via : Initial Calibration

